

NIST focuses on testing standards to support lab on a chip commercialization

August 9 2012



This is a microfluidic lab on a chip device sitting on a polystyrene dish. Stainless steel needles inserted into the apparatus serve as access points for fluids into small channels within it. Credit: Gregory A. Cooksey/NIST

Lab on a chip (LOC) devices -- microchip-size systems that can prepare and analyze tiny fluid samples with volumes ranging from a few microliters (millionth of a liter) to sub-nanoliters (less than a billionth of a liter) -- are envisioned to one day revolutionize how laboratory tasks such as diagnosing diseases and investigating forensic evidence are performed. However, a recent paper* from the National Institute of Standards and Technology (NIST) argues that before LOC technology can be fully commercialized, testing standards need to be developed and implemented.

"A testing standard," explains NIST physical scientist and paper author

Samuel Stavis, "defines the procedures used to determine if a lab on a [chip](#) device, and the materials from which it is made, conform to specifications." [Standardized testing](#) and measurement methods, Stavis writes, will enable MEMS ([microelectromechanical systems](#)) LOC manufacturers at all stages of production—from processing of raw materials to final rollout of products—to accurately determine important physical characteristics of LOC devices such as dimensions, electrical surface properties, and fluid flow rates and temperatures.

To make his case for testing standards, Stavis focuses on autofluorescence, the background fluorescent glow of an LOC device that can interfere with the analysis of a sample. Stavis states that multiple factors must be considered in the development of a testing standard for autofluorescence, including: the materials used in the device, the measurement methods used to test the device and how the measurements are interpreted. "All of these factors must be rigorously controlled for, or appropriately excluded from, a meaningful measurement of autofluorescence," Stavis writes.

Quality control during LOC device manufacturing, Stavis says, may require different tests of autofluorescence throughout the process. "There may be one measure of autofluorescence from the block of plastic that is the base material for a chip, another once the block has been fashioned into the substrate in which the functional components are embedded, and yet another as the final device is completed," Stavis says. "To manufacture lab on a chip devices with reliably low autofluorescence, accurate measurements may be needed at each stage."

Stavis also emphasizes that it is important not to confuse testing standards with product standards, and to understand how the former facilitates the latter. "A product standard specifies the technical requirements for a lab on a chip [device](#) to be rated as top quality," he says. "A testing standard is needed to measure those specifications, as

well as to make fair comparisons between competing products."

Provided by National Institute of Standards and Technology

Citation: NIST focuses on testing standards to support lab on a chip commercialization (2012, August 9) retrieved 3 May 2024 from <https://phys.org/news/2012-08-nist-focuses-standards-lab-chip.html>

| |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <p>This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.</p> |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|