


# Intracellular recordings using nanotower electrodes

April 8 2016

---



120- $\mu\text{m}$ -height 'nanotower' electrode is punching a cell membrane. Silicon growth technology and three-dimensional nano/microfabrication techniques realize such high-aspect-ratio intracellular electrodes. Credit: (c) Toyohashi University Of Technology.

Our current understanding of how the brain works is very poor. The electrical signals travel around the brain and throughout the body, and the electrical properties of the biological tissues are studied using electrophysiology. For acquiring a large amplitude and a high quality of neuronal signals, intracellular recording is a powerful methodology compared to extracellular recording to measure the voltage or current across the cell membranes. Nanowire- and nanotube-based devices have been developed for the intracellular recording applications to demonstrate the advantages of these devices having high spatial resolution and high sensitivity.

However, length of these nanowire/nanotube electrode devices is currently limited to less than 10  $\mu\text{m}$  due to process issues that occur during fabrication of high-aspect-ratio nanoscale devices, which are more than 10- $\mu\text{m}$  long. Thus, conventional nanodevices are not applicable to neurons/cells within thick [biological tissues](#), including brain slices and brain in vivo.

A research team in the Department of Electrical and Electronic Information Engineering and the Electronics-Inspired Interdisciplinary Research Institute (EIIRIS) at Toyohashi University of Technology has developed three-dimensional microneedle-based nanoscale-tipped electrodes (NTEs) that are longer than 100  $\mu\text{m}$ . The needle length exceeds that of the conventional nanowire/nanotube-based intracellular devices, thus expanding the range of applications of nanodevices in intracellular recording, such as deep tissue penetration. Additionally, they perform intracellular recordings using muscle cells.

"A technological challenge in electrophysiology is intracellular recordings within a thick biological tissue. For example, a needle length of more than 40  $\mu\text{m}$  is necessary for performing brain slice experiments. However, it is almost impossible to penetrate nanoscale diameter needles with a high-aspect-ratio, because of the long hair-like nanostructure that

has insufficient stiffness. On the other hand, our NTE, which is 120- $\mu\text{m}$ -long cone-shaped electrode, has sufficient stiffness to punch tissues and cells", explains the first author PhD candidate, Yoshihiro Kubota.

The leader of the research team, Associate Professor Takeshi Kawano said "Although we demonstrated the preliminary results of our NTE device, the batch fabrication of such intracellular electrodes, which have a needle length more than 100  $\mu\text{m}$ , should lead to an advancement in the device technologies. This will eventually lead to realization of multisite, depth-intracellular recordings for biological tissues, including brain slices and brain in vivo, which are beyond the capability of conventional intracellular devices."

As addressed by the research team, the NTE has the potential to be used in cells that are deep within a biological tissue, including brain slice and brain in vivo, thus accelerating the understanding of the [brain](#).

**More information:** Yoshihiro Kubota, Hideo Oi, Hirohito Sawahata, Akihiro Goryu, Yoriko Ando, Rika Numano, Makoto Ishida, and Takeshi Kawano (2016). Nanoscale-tipped high-aspect-ratio vertical microneedle electrodes for intracellular recordings, *Small*, Article first published online: 8 April 2016 | [DOI: 10.1002/sml.201600172](https://doi.org/10.1002/sml.201600172)

Provided by Toyohashi University of Technology

Citation: Intracellular recordings using nanotower electrodes (2016, April 8) retrieved 4 May 2024 from <https://phys.org/news/2016-04-intracellular-nanotower-electrodes.html>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.
---