


# On-chip pumps achieve high-speed sorting of large cells

July 28 2017



The photograph shows the developed microfluidic chip which enables us to sort cells at high speed of 16 microseconds. The enlarged view shows a demonstration of on-chip cell sorting of a *Euglena gracilis* cell (time interval of each frame: 40  $\mu$ s). © Shinya Sakuma, Yusuke Kasai, Takeshi Hayakawa, and Fumihito Arai . Credit: Nagoya University

A research group in Nagoya University developed a high-speed cell sorting method of large cells with high-viability using dual on-chip

pumps.

The sorting of individual cells is necessary for many biological applications, including the isolation of specific cell types from cell suspensions. A fluorescence-activated cell sorting (FACS) has been used for high-throughput cell sorting. In this method, lasers are used to excite auto-fluorescence or tagged-fluorescence of cell included in droplets, and then droplets are diverted into different containers depending on their characteristics. However, this technique is concerned about sample infections due to aerosols generation. Additionally, a FACS of larger cells requires the samples to be processed under low pressure through wider nozzles to prevent damage. Thus, sorting is limited to low-level throughput.

Research at Nagoya University on cell sorting used a microfluidic chip to prevent sample infection. This chip has microchannels into which cell suspensions are introduced for sorting. The research group integrated two externally-driven on-chip pumps into the [microfluidic chip](#) for high-speed flow control. Using a high-speed actuator as the driving source of pump, they succeeded in producing a flow with 16 microseconds for cell sorting.

Microfluidic chip contains a cross-shaped sorting area and three-branched microfluidic channel. "Target/non-target cells are three-dimensionally aligned in the main channel," corresponding author Shinya Sakuma says. "When target cells are detected, the on-chip pumps work rapidly to sort cells into one of two interest channels. Meanwhile, non-target cells are flushed into the waste channel without pump actuation."

The technique allows us to sort not only large but also small cells with high speed, high purity, and high viability. "We tested the method on microalgae as an example of large [cells](#), around 100 micrometers in size, and achieved 95.8 percent purity, 90.8 percent viability, and a 92.8

percent success rate," corresponding co-author Yusuke Kasai says. "As a model small cell type, we used a cancer cell whose size is around 24 micrometers, and achieved 98.9 percent purity, 90.7 percent viability, and a 97.8 percent success rate."

**More information:** Shinya Sakuma et al. On-chip cell sorting by high-speed local-flow control using dual membrane pumps, *Lab Chip* (2017). DOI: [10.1039/c7lc00536a](https://doi.org/10.1039/c7lc00536a)

Provided by Nagoya University

Citation: On-chip pumps achieve high-speed sorting of large cells (2017, July 28) retrieved 9 April 2024 from <https://phys.org/news/2017-07-on-chip-high-speed-large-cells.html>

<p>This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.</p>
--