

Piezoelectric and laser ultrasonic system takes 3-D ultrasound images of solids

September 25 2020



High-resolution 3-D imaging result of branched stress corrosion cracking. Credit: Yoshikazu Ohara, Tohoku University

A new system, developed by Tohoku University researchers in Japan in collaboration with Los Alamos National Laboratory in the US, takes 3-D



images that can detect defects in metallic structures. The approach was published in the journal *Applied Physics Letters* and could enhance safety in power plants and airplanes.

Yoshikazu Ohara and colleagues at Tohoku University use nondestructive techniques to study structures, and wanted to find a way to produce 3-D images of structural defects. They developed a new technology, called the piezoelectric and laser ultrasonic system (PLUS), that combines the strengths of two different devices to produce highresolution 3-D images of defects in metallic structures

"We believe that PLUS will pave the way for accurate evaluation of material strength, the identification of defects, and finding out how defects initially started to form," says Ohara.

Currently available "ultrasonic phased arrays" are a powerful tool for imaging internal defects in solids, but only in two dimensions. These devices are made of a piezoelectric one-dimensional array transducer with a limited number of individual elements—up to 128. Electrical pulses in the piezoelectric elements are converted to a mechanical vibration that emits <u>ultrasonic waves</u> into the material under investigation. Ultrasonic waves are reflected back from internal defects and converted into electric signals that can be translated into a 2-D image.





A Schematic illustration of the high-resolution 3D ultrasonic imaging system (PLUS). Credit: Yoshikazu Ohara, Tohoku University

In PLUS, the waves generated in a material from a <u>piezoelectric</u> <u>transducer</u> with a single element are received by a laser Doppler vibrometer, which moves around the material's surface to get a good 2-D scan of the area. As a result of this process, it receives the scattered and reflected waves at a much larger number of "points" than those that can be received by a piezoelectric array transducer. The information received by the laser Doppler vibrometer is transmitted by an oscilloscope to a computer, where it is processed by an imaging algorithm and converted into a 3-D image.

"Ultrasonic phased arrays, which are on the cutting-edge of ultrasonic



inspection, can only provide 2-D images because of their limited number of elements," says Ohara. "PLUS makes it possible to have thousands of elements as a result of incorporating the 2-D scan of a laser Doppler vibrometer in place of a piezoelectric array transducer."

Although tested only on defects in metallic materials, Ohara says their technology can be applied to other materials, including concrete and rock, simply by changing the phased array transmitter to one that emits a different range of ultrasound frequencies.

One drawback is the long data acquisition and processing time, which takes several hours. However, this can be shortened by adopting a high-speed analog-to-<u>digital converter</u> in place of the oscilloscope, using a more sensitive <u>laser</u> Doppler vibrometer, utilizing different imaging algorithms, and employing a graphical processing unit.

More information: Yoshikazu Ohara et al. Toward an ultra-high resolution phased-array system for 3D ultrasonic imaging of solids, *Applied Physics Letters* (2020). DOI: 10.1063/5.0021282

Provided by Tohoku University

Citation: Piezoelectric and laser ultrasonic system takes 3-D ultrasound images of solids (2020, September 25) retrieved 16 June 2024 from <u>https://phys.org/news/2020-09-piezoelectric-laser-ultrasonic-d-ultrasound.html</u>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.