

An improved safety standard for bionic devices

March 29 2021



Credit: CC0 Public Domain

Applied physicists at the University of Sydney have proposed new standards to measure moisture leaks into bionic devices such as pacemakers, cochlear hearing implants and retinal replacements.

The researchers, who received an industry partnership funding through the Australian Research Council to undertake the study, say the new moisture standards could give the wearers of bionic implants extra confidence in the operation of the life-changing devices. They also say that the improved moisture-testing regime could be used in the emerging renewable energy industry where new-generation solar cells require high standards of humidity control.

Bionic implants must be able to operate successfully in moist environments in the human body. While the potential for large leaks into the devices are easy to detect during manufacturing, small leaks can escape detection and standard testing is required to ensure safety and prevent moisture-induced failure.

Professor David McKenzie from the School of Physics at the University of Sydney said: "The [accurate measurement](#) of moisture penetration into medical devices is essential to guarantee long-term performance. Accurate measurement needs an accurate industry standard to assess leak risks."

He said there are commercially available systems that measure relative humidity, but these are not sensitive enough for the most demanding applications in implantable biomedical devices. Using mass spectroscopy technology, the measurement of helium as a proxy for moisture leakage is a de facto industry standard test for the critical small leaks that are hard to pick up.

In practice and in most cases, helium testing of [bionic devices](#) is a good standard, but by improving the compliance by a factor of 10, we think the industry can further guarantee the safety of biomedical implants," Professor McKenzie said.

"Helium leak detection tests are widely used for the location of more

sensitive leaks, but direct measurement of gaseous or liquid water leakage at similar sensitivities has proven too difficult until now," he said.

Professor McKenzie and colleagues from the Australian government's National Measurement Institute and have developed a standard leak test and shown how to use it to verify the quality of encapsulations, or the containment of medical devices.

They have published their research in the American Chemical Society's *Applied Materials & Interfaces*.

Lead author Dr. Wenwen Lei from the National Measurement Institute said: "For very high sensitivities, it is important we try to directly measure water leak risks rather than the de facto measure of helium, which is not as accurate.

"Water has special abilities to penetrate through very small spaces given its different behavior. For instance, it has been recently discovered that water acts in a really weird way in carbon nanotubes, where the water does not seem to touch the sides of tiny passageways not much wider than a millionth of a human hair."

Professor McKenzie said: "Any such 'nanoleakage' would be really serious if it happened in medical devices, especially those with lots of feedthrough apertures, such as the bionic eye."

The new standard moisture leak breaks new ground by studying water behavior in a stable polymer material. Using the new tests, researchers directly compared water leakage to helium leakage in biomedical test structures. They found that in the majority of cases, calibrated measurements of moisture leak rates do exceed helium leak rates, especially for very small and large leaks.

Given their findings that [water](#) can [leak](#) up to 10 times the rate of helium leaks, they propose strengthening the compliance standard for [helium](#) testing of biomedical devices tenfold, or one order of magnitude.

"As we improve bionic technology, wearable implants will become more common, so it's important to give people that extra level of safety assurance," Professor McKenzie said.

More information: Wenwen Lei et al, Quantifying Moisture Penetration in Encapsulated Devices by Heavy Water Mass Spectrometry: A Standard Moisture Leak Using Poly(ether-ether-ketone), *ACS Applied Materials & Interfaces* (2021). [DOI: 10.1021/acsami.0c23115](#)

Provided by University of Sydney

Citation: An improved safety standard for bionic devices (2021, March 29) retrieved 16 June 2024 from <https://phys.org/news/2021-03-safety-standard-bionic-devices.html>

<p>This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.</p>
--