

Developing multiple concentration gradients for single cell–level drug screening

May 1 2023, by Thamarasee Jeewandara



Drug screening at the single-cell level based on multiconcentration gradient construction and a single-cell capture device. (a) Schematic diagram of the integrated microfluidic device. (b) The detailed design of a single-cell capture device. The various channels are shown with different colors to visualize the



microfluidic device's different components. Red and green indicate the fluidic channels of cells and reagents, respectively, and blue shows the control channels and valves. (c) Schematic of manipulation of single-cell capture according to cell size and deformability. The procedure consists of three steps: cell infusion, single-cell capture, and reagent infusion. Credit: *Microsystems & Nanoengineering* (2023). DOI: 10.1038/s41378-023-00516-0

The potential to isolate and regulate the biodynamics of single cells is significant in drug design and screening. However, pre-existing experimental reports in single-cell drug screening must yet provide multiple-dose gradient studies, to accurately predict drug–single cell interactions and performance. In a new report in *Microsystems and Nanoengineering*, Shaofei Shen and a team of scientists in life sciences and medicine in China solved this problem by creating a multi-concentration gradient generator.

They combined the gradient generator with a single-cell capture array, to understand the impact of single or combined doses of 5-fluoro uracil and cisplatin anticancer drugs on human hepatoma cells and human breast cancer cells, at the single-cell level. The instrument provided a simple and reliable platform to study the correct dosage of different drug candidates at the single cell level to screen single-agent chemotherapy agents and efficiently form combinatorial therapy regimes.

Developing a new drug screening platform

Drug screening methods offer a visible solution to prevent infections and <u>treat human disease</u>. Broad research efforts have shown that microfluidic chip technology offers a microanalysis platform for <u>easy</u> <u>access and biocompatibility</u>. Most <u>microfluidic systems</u> offer a powerful instrument to study cell populations at the level of single cells, while



facilitating dose-dependent cellular responses at different drug concentrations. In this work, Shen and colleagues optimized a multistage microfluidic device combined with a single-cell capture array to generate a single-cell microfluidic drug screening platform across multiple drug doses.



Generation of microfluidic drug concentration gradient under different flow conditions. (a) Simulation imaging of the generation process of the drug concentration gradient in the device. Three inlets were filled with drug A, drug B, and drug C when the flow rate was 10 μ L/min. (b) Numerical simulation of drug concentration gradient formation in 24 microchambers. (c) Concentration gradients formed by fluorescence experiments in 24 microchambers. The



transport and distribution of substances in the system and the formation of a concentration gradient were fluorescence characterized using luciferin as a model drug. Scale bar: 1200 μ m. (d) Fluorescence experiment and computer-simulated concentration value in 24 microchambers. The error bars refer to the standard deviations obtained from ten parallel experiments. (e) Linear relationship of drug concentration in the first nine microchambers by fluorescence experiment and numerical simulation under different flow conditions. Credit: *Microsystems & Nanoengineering* (2023). DOI: 10.1038/s41378-023-00516-0

In this work, the researchers calculated the <u>concentration gradient</u> generated in theory, and verified it in a fluorescent experiment. They used <u>cisplatin</u> and <u>5-fluorouracil</u>; two chemotherapeutic agents as model drugs for single or multidrug combinatorial chemotherapy on <u>human</u> <u>breast cell carcinoma cells</u> and <u>human hepatoma cells</u> at the single-cell level. The system provided a flexible and well-regulated instrument to study pharmacological functions and conduct single-cell research.

Designing the microfluidic platform

Shen and the team designed and constructed the microfluidic platform containing 24 single-cell capture devices. They generated a series of successful drug concentration gradients in the device, where the concentration-gradient generator could jointly screen two drugs, allowing them to study a single drug and multidrug combinations alongside <u>their optimal dosage</u>.

The bioengineers facilitated the single-cell capture device to contain a two-dimensional array with cell and reagent portals for cell suspension and reagent import, as <u>described in a previous study</u> by the same team.



Tai Chi-spiral mixer–induced Dean flow. (a) Sketch diagram of spiral mixer channels. (b–d) Dean flow simulation at different positions in the channel when the flow rate was 10 μ L/min (b), 20 μ L/min (c), and 50 μ L/min (e). Scale bar: 10 μ m. e Quantitative analysis of cross-sectional Dean flow. The results are obtained at the positions shown in the dotted lines of Fig. 3b–d. Credit: *Microsystems & Nanoengineering* (2023). DOI: 10.1038/s41378-023-00516-0

Three-concentration gradient formation

The researchers tested the capacity of the device to establish three concentration gradients via <u>numerical simulations</u> and fluorescein



experiments, to explore drug distributions in 24 microcavities. Using an inverted microscope, they gathered fluorescent images and used imaging software to understand the data gathered for each chamber.

The simulation showed the formation of three groups of identical drug concentrations in the designed device and their percentage. Alongside the simulations, they conducted fluorescein experiments to verify the concentration gradient distribution in 24 liquid storage chambers. They injected two fluorescent agents into the chip from the entrances and observed how the experimental outcomes agreed with the results of numerical simulations.

a	С18-0µМ&12.5µМ	G 19 0µМ825µ М	С20-0µ1М837.5µ1М	C21-01M8501M	С22-0µМ&62.5µМ	С23-0µМ&75µМ
DDP&5-FU	С24-0µМ&87.5µМ	С1-оµМ&100µМ	С16-1.25µМ80µМ	Q15-2.5µM&0µM	С14-3.75µM80µM	С13-5µМ80µМ
	С12-6.25µМ80µМ	C11-7.5µM&0µM	С 10-8.75µМ80µМ	Q9-10µМ&ОрМ	С17-0µМ&0µМ	C8-8.75µM812.5µM
	G7-7.5µM&25µM	С 6-6 .25µМ&37.5µМ	С5-5µМ850µМ	C4-3.75µM&62.5µM	С3-2.5µМ&75µМ	C2-1.25µM887.5µM
b	С18-0µМ&12,5µМ	С19-0µМ&25µМ	С20-0µМ&37.5µМ	сёт-орм&зони	С22:0µМ#62:5µМ	С23-0µМА75µМ
5-FU	C24-0µM887.5µM	63-0µ.M81000µ.M	С16-1.25µМ&0µМ	С15-2.5µМ&0µМ	С143.75µМ80µМ	С13-5µМВФµМ
DDP&	С12-6.25µМ&0µМ	С11-7.5µМ&0µМ	С10-8.75µМ80µМ	С9-10µМ80µМ	С17-0µМ&0µМ	CB-8.75mM&12.5mM
	С7-7.5µМ&25µМ	06-6.25µМ &37.5µМ	С5-5µМ&50µМ	С4-75µМ&62.5µМ	С3-2.5µМ&75µМ	С2-1.25µМ&87.5µМ



Response of tumor cells in the largest filter units of single-cell capture structures to multiple-gradient dosages of two drugs (5-FU, DDP). (a) Fluorescence images of HepG2 cells were obtained by AO/PI staining after continuous treatment with different concentrations of drugs for 2 h. (b) Fluorescence images of MCF-7 cells were obtained by acridine AO/PI staining after continuous treatment with different concentrations of drugs for 2 h. Scale bar: 800 µm Credit: *Microsystems & Nanoengineering* (2023). DOI: 10.1038/s41378-023-00516-0

Tai Chi-spiral mixer-induced Dean flow in microfluidics

The team used a Tai Chi-spiral mixer in the concentration gradient generator and incorporated <u>Dean flow in the curving channels</u> for microfluidic applications of fluid regulation, as a revolutionary approach to mix, purify and focus on <u>reactions efficiently and cost-effectively</u>.

They obtained a stable and effective multidrug concentration gradient for single-cell drug screening via the formation of three concentration gradients and constructed a multiple-concentration platform for singlecell level drug screening to observe the combined and separate effects of the anticancer drugs cisplatin and 5-fluorouracil.

Drug screening with single-cell-drug interactions

The bioengineers applied a drug containing medium with different concentrations of the two cell types; human breast cell carcinoma cells and human hepatoma cells in single-cell capture microfluidic devices and in traditional Petri dishes. Using double fluorescence staining methods; <u>acridine orange</u> and <u>propidium iodide</u>, they validated the cell viability and observed variations in cell activity when exposed to



different drug concentrations to show how the survival rate of the tumor cells improved as the concentration of the drug decreased in the culture devices.

While the cell vitality negatively correlated with the drug dose in healthy cells, tumor cells were unaffected by cell interactions at the single-cell level for <u>effective investigations of their susceptibility to drugs</u>.





Quantitative comparison of cell viability under multiple-gradient dosages of two drugs (5-FU, DDP). (a, b) Quantitative comparison of MCF-7 cell viability (a) and HepG2 cell viability (b) in single-cell level culture in microarray and Petri dishes. (c) d Quantitative comparison of MCF-7 cell viability (c) and HepG2 cell viability (d) in the different biomechanical heterogeneity of tumor cells at the single-cell level. Credit: *Microsystems & Nanoengineering* (2023). DOI: 10.1038/s41378-023-00516-0

The team combined two drugs on the cells in the platform and noted a stronger synergistic effect of two drugs compared to monotherapy alone. The phenomenon is also present in microfluidic single-cell capture devices to offer effective monotherapy and combination therapy strategies. For instance, cells exposed to 5-fluorouracil underwent dose lethality to inhibit DNA synthesis and cell mitosis, highlighting the significance of drug screening to explore tumor cell heterogeneity across diverse drug gradients. The work has significant impact on broader biological and preclinical explorations, including cancer stem cell separation and drug discovery.

Outlook

In this way, Shaofei Shen and colleagues generated a simple and efficient multifunctional microfluidic drug screening device for singlecell level functionality. The device contained a concentration <u>gradient</u> drug generator and single-cell capture array suited for multiple purposes.

They combined the single-cell capture device on the platform to implement separate doses of two <u>anticancer drugs</u>, on two different cancer cell lines. The work provides a strong starting point to study the sensitivity of multiple anticancer agents at the level of single cells and



stem cells, to implement effective chemotherapeutic strategies.

More information: Shaofei Shen et al, Construction of multiple concentration gradients for single-cell level drug screening, *Microsystems & Nanoengineering* (2023). DOI: 10.1038/s41378-023-00516-0

Leonard I. Zon et al, In vivo drug discovery in the zebrafish, *Nature Reviews Drug Discovery* (2005). DOI: 10.1038/nrd1606

© 2023 Science X Network

Citation: Developing multiple concentration gradients for single cell–level drug screening (2023, May 1) retrieved 25 June 2024 from <u>https://phys.org/news/2023-05-multiple-gradients-celllevel-drug-screening.html</u>

This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.