


Researchers present simplified, improved scheme for precision measurement using lasers

January 15 2024, by Kenna Hughes-Castleberry



A schematic of a laser going through an AOM, which sends sound waves into a silicon cavity. Credit: Kenna Hughes-Castleberry/JILA/Ye and Hall Groups

Within atomic and laser physics communities, scientist John "Jan" Hall has become a key figure in the history of laser frequency stabilization and precision measurement using lasers. Hall's work revolved around understanding and manipulating stable lasers in ways that were revolutionary for their time. His work laid a technical foundation for measuring a tiny fractional distance change brought by a passing gravitational wave. His work in laser arrays awarded him the Nobel Prize in Physics [in 2005](#).

Building on this foundation, JILA and NIST Fellow Jun Ye and his team embarked on an ambitious journey to push the boundaries of precision measurement even further. This time, their focus turned to a specialized technique known as the Pound-Drever-Hall (PDH) method (developed by scientists R. V. Pound, Ronald Drever, and Hall himself), which plays a large role in precision optical interferometry and laser [frequency](#) stabilization.

While physicists have used the PDH method for decades in ensuring their laser frequency is stably "locked" to an artificial or quantum reference, a limitation arising from the frequency modulation process itself, called residual amplitude modulation (RAM), can still affect the stability and accuracy of the laser's measurements.

In a recent [Optica](#) paper, Ye's team, working with JILA electronic staff member Ivan Ryger and Hall, implemented a new approach for the PDH method, reducing RAM to never-before-seen minimal levels while

making the system more robust and simpler.

As the PDH technique is implemented in various experiments, from gravitational wave interferometers to optical clocks, improving it further offers advancements to a range of scientific fields.

A dive into laser 'locking'

Since its publication [in 1983](#), the PDH method has been cited and utilized thousands of times. "Setting up a PDH lock is something you might learn in an undergraduate lab course; that's just how central it is doing all the experiments we do in [atomic physics](#)," explained recently awarded Ph.D. candidate Dhruv Kedar, the paper's co-first author.

The PDH method uses a frequency modulation approach to precisely measure the laser frequency or phase fluctuations. The frequency modulation adds special "sidebands" (or additional light signals) around a main light beam, known as the "carrier."

Comparison of these sidebands against the main carrier helps measure any slight changes in the frequency or phase of the main light beam relative to a reference. This technique is especially useful because it's very sensitive and can reject unwanted noise and errors.

Physicists can then use these combined light beams to interrogate different environments, such as an optical cavity made of mirrors. To do this, the researchers must "lock" the laser to the cavity or have it probe the cavity at a particular frequency.

"What that means is that you're trying to lock your laser to the center of your resonance," Kedar added. This allows the laser to reach state-of-the-art levels of stability, which is especially important when trying to tease out tiny changes in the optical length or monitoring quantum dynamics,

such as energy shifts or spin changes in atoms and molecules.

Unfortunately, "locking" a laser doesn't always mean it stays stable or "in resonance with the center of the optical cavity, as noise like RAM can change the relative offsets of the reference light beams and introduces frequency shift," co-first author and JILA Postdoc Zhibin Yao elaborated. "The RAM can contaminate your PDH error signal."

As the JILA researchers quickly realized, along with the rest of the laser physics community, reducing this RAM is crucial for improving the stability of the PDH technique and, in turn, their laser measurements. Overcoming the RAM problem has been a long journey, but the new approach would make the fight much easier.

Reducing RAM via EOMs and AOMs

The two-reference light "sidebands" are essential for the PDH locking method. To generate the "sidebands," the JILA researchers needed to use a frequency modulator, either an Electro-Optic Modulator (EOM) or an Acousto-Optic Modulator (AOM).

Historically, EOMs have been employed in various optical systems by applying electric fields to optical crystals to change the phase of laser light coming through the crystal. When an electric field is applied to certain types of crystals, it modulates the laser phase by altering the crystal's refractive index. This process allows EOMs to add sidebands to the carrier beam easily.

However, the effective phase modulation of the crystal used in EOMs is easily altered by environmental fluctuations, introducing RAM into the PDH error signal and consequently making it less stable. In contexts where ultra-high precision is required, such as running an optical timescale or operating an atomic clock, even minuscule amounts of

RAM can introduce fluctuations at undesired levels.

"EOMs add sidebands to the carrier laser in the optical domain, which is more challenging for us to control," Kedar explained. "So instead, we can try to generate these sidebands in the electronic domain and translate them to the optical by using an AOM."

AOMs represent a newer approach to reducing RAM by using [sound waves](#) to modulate the laser light. When a sound wave propagates through a crystal or a transparent medium, it creates a diffraction pattern that bends the laser light in various amounts. As a light beam passes through this sound wave-altered medium, the variations in refractive index act like a series of tiny prisms, altering the path and, thus, the frequency of the light.

Kedar added, "If you want to control the amplitude of each sideband, you control the amplitude of the main tone that you're generating in the microwave domain via the AOM." Because the AOM doesn't modulate the laser frequency based on the electro-optic effect, it produces much less RAM noise than EOM, reducing the overall RAM level of the system. All of the beams coming out of the AOM crystal can be combined in a single optical fiber, putting all frequency shift beams into a single, common spatial mode profile.

Comparing EOM and AOM

To measure the advantages of this new PDH approach, Kedar, Yao, Ye, and the rest of the team ran an experiment using the traditional EOM and their improved AOM setup and compared the results. They found that with the AOM, they could reduce the RAM levels to a small fraction of parts per million. Equally importantly, this approach allows much more flexibility in controlling relative strength between the carrier and two sidebands. The AOM advantage is much more obvious when the

carrier becomes vanishingly small.

"Instead of parts per million, you can do like 0.2 parts per million, which seems like a small improvement, but that's kind of toeing the line for acceptable levels of RAM for us," Kedar said. "Even though this RAM level is so small, it's still a significant roadblock to improving our cavities and making them slightly better. That extra factor of two or three is enormously helpful in pushing the frontiers of state-of-the-art [laser](#) stabilization."

The simple implementation of AOM instead of EOM suggests an answer even Hall would be proud of. "It's simple enough that, in principle, someone can look at this scheme and see it as a natural method to interrogate a spectral feature," Kedar said. "In the end, this speaks to the research style that Jan and Jun both create: a very elegant, simple solution."

More information: Dhruv Kedar et al, Synthetic FM triplet for AM-free precision laser stabilization and spectroscopy, *Optica* (2023). [DOI: 10.1364/OPTICA.507655](https://doi.org/10.1364/OPTICA.507655)

Provided by JILA

Citation: Researchers present simplified, improved scheme for precision measurement using lasers (2024, January 15) retrieved 12 May 2024 from <https://phys.org/news/2024-01-scheme-precision-lasers.html>

<p>This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.</p>
--