

Highly sensitive diamond quantum magnetometer can achieve practical ambient condition magnetoencephalography

June 6 2024

High-Sensitivity Diamond Quantum Magnetometer For Ambient Magnetoencephalography

Diamond quantum magnetometers are promising for realizing ambient condition magnetoencephalography (MEG)



However, diamond quantum magnetometers with sufficient sensitivity for practical MEG have not yet been developed



Continuous-wave optically detected magnetic resonance-based diamond magnetometer



Credit: *Physical Review Applied* (2024). DOI: 10.1103/PhysRevApplied.21.064010

A highly sensitive diamond quantum magnetometer utilizing nitrogen-



vacancy centers can achieve millimeter-scale resolution magnetoencephalography (MEG), as reported by scientists from Tokyo Tech. The novel magnetometer, based on continuous-wave optically detected magnetic resonance, marks a significant step towards realizing ambient condition MEG and other practical applications.

MEG is a biomedical imaging technique used for mapping brain activity by recording magnetic fields produced by the naturally occurring electrical currents generated by neurons in the brain, using very sensitive magnetometers.

Currently, MEG requires a magnetically shielded room for operation. Achieving MEG that works in normal environments, without the need for magnetic shielding, is a major goal. This would enable daily diagnosis, <u>brain-machine interfaces</u>, and fundamental research on brain function.

Magnetometers using diamond quantum sensors with nitrogen–vacancy (NV) centers are promising candidates for realizing ambient condition MEG. These sensors are expected to offer significantly better millimeter-scale resolution than conventional centimeter-scale MEGs. Notably, NV centers are defects in the structure of a diamond, consisting of a nitrogen atom substituted for a carbon atom, next to a vacancy.

One common method for measuring magnetic fields with NV centers is continuous-wave optically detected <u>magnetic resonance</u> (CW-ODMR). In this method, a continuous microwave field is used to manipulate the spin states of the NV centers while they are illuminated by a laser. The intensity of this laser-induced fluorescence changes depending on the external magnetic field.

By measuring these changes in fluorescence, the external magnetic field can be detected and measured. Compared with other methods, this



method is simpler and easier and can achieve millimeter-scale resolution.

Building upon this technology, a team of researchers from Japan, led by Associate Professor Naota Sekiguchi from the Department of Electrical and Electronic Engineering at Tokyo Institute of Technology, recently developed a novel and sensitive diamond quantum <u>magnetometer</u>.

"Clinically acceptable MEG requires sensitivities, at least, in the order of picotesla (pT) in the near-dc frequency range of 5–100 Hz within a reasonable measurement time. Currently, diamond magnetometers require a magnetic flux concentrator (MFC) to achieve this sensitivity. However, MFCs reduce the intrinsic spatial resolution of magnetometers.

"Additionally, for MEGs, a short measurement distance is required as the decay of magnetic fields increases exponentially with distance. To overcome these challenges, we designed a sensitive CW-ODMR-based diamond magnetometer," explains Sekiguchi.

Their study has been <u>published</u> in the journal *Physical Review Applied*.

The novel magnetometer uses a single crystalline diamond which was fabricated using a high-pressure high temperature (HPHT) method. After HPHT synthesis, a piece of crystal was cut out parallel to the (111) crystal plane, and negatively charged NV centers were produced in the crystal using electron beam irradiation followed by annealing at 1,000 °C.

This NV center ensemble was placed in the sensor head, designed to approach the target to about one millimeter with a sensing volume of 4 x 10^{-3} mm³. The ensemble was excited by a linearly polarized green laser with a wavelength of 532 nanometers, and a high refractive index hemispherical lens was used to enhance the collection efficiency of the



laser-induced fluorescence.

By carefully tuning the experimental conditions, the researchers achieved a record sensitivity of 9.4 ± 0.1 pT Hz^{-1/2} in the frequency range of 5–100 Hz, without an MFC. Additionally, analysis of Allan deviation showed that the magnetometer can measure magnetic fields as low as 0.3 pT and maintain remarkable sensitivity for a long time. Furthermore, its design is suitable for practical applications such as for MEG of a living animal.

The high sensitivities achieved in this study mark a significant step toward realizing ambient condition MEG with millimeter-scale resolution. Looking ahead, Sekiguchi concludes, "In the future, we plan to measure the MEG of animals using the sensors developed in this study and to realize MEG measurements with diamond quantum sensors. Ultimately, we aim to achieve MEG without the need for magnetic shielding."

More information: N. Sekiguchi et al, Diamond quantum magnetometer with dc sensitivity of sub-10 pT Hz–1/2 toward measurement of biomagnetic field, *Physical Review Applied* (2024). DOI: 10.1103/PhysRevApplied.21.064010. On *arXiv*: DOI: 10.48550/arxiv.2309.04093

Provided by Tokyo Institute of Technology

Citation: Highly sensitive diamond quantum magnetometer can achieve practical ambient condition magnetoencephalography (2024, June 6) retrieved 23 June 2024 from <u>https://phys.org/news/2024-06-highly-sensitive-diamond-quantum-magnetometer.html</u>



This document is subject to copyright. Apart from any fair dealing for the purpose of private study or research, no part may be reproduced without the written permission. The content is provided for information purposes only.